

HUMAN
B CELL
ENRICHMENT
KIT

CATALOG #19054VX

THIS PRODUCT INFORMATION SHEET IS PROVIDED FOR USE WITH THE "EASY 50" EASYSEP® MAGNET (SECTION A), THE PURPLE EASYSEP® MAGNET (SECTION B) OR "THE BIG EASY" SILVER EASYSEP® MAGNET (SECTION C).**A) MANUAL PROTOCOL USING THE "EASY 50" EASYSEP® MAGNET (CATALOG #18002)**This procedure is used for processing **1 mL to 40 mL** of sample (up to 2×10^9 cells).1. Prepare nucleated cell suspension at a concentration of 5×10^7 cells/mL in recommended medium (see Notes and Tips, reverse side). Cells must be placed in a 50 mL conical tube to properly fit into the Easy 50 EasySep® magnet.*Falcon™ 50 mL Polypropylene Conical Tubes (BD Biosciences, Catalog #352070) are recommended.*2. Add EasySep® Human B Cell Enrichment Cocktail at **50 µL/mL cells** (e.g. for 2 mL of cells, add 100 µL of cocktail). Mix well and incubate at room temperature (15 - 25°C) for **10 minutes**.

3. Vortex EasySep® D Magnetic Particles for 30 seconds. Ensure that the particles are in a uniform suspension with no visible aggregates.

4. Add the EasySep® D Magnetic Particles at **75 µL/mL** of cells (e.g. for 2 mL of cells, add 150 µL of particles). Mix well and incubate at room temperature for **10 minutes**.5. Add the recommended medium to bring the cell suspension to a **total volume** according to the table below:

ORIGINAL SAMPLE VOLUME	1-10 mL	10-40 mL
TOTAL VOLUME	25 mL	50 mL

6. Mix the cells in the tube by gently pipetting up and down 2-3 times. Place the tube (without the cap) into the magnet and push all the way down. Set aside for **10 minutes**.

7. Carefully remove the enriched cell suspension by pipetting off into a new 50 mL tube. Do not pour. Ensure that the pipette does not touch the sides of the tube or bottom of the tube. The magnetically labeled unwanted cells will remain bound inside the tube, held by the magnetic field of the magnet. The negatively selected enriched cells, in the new tube are now ready for use.

B) MANUAL EASYSEP® PROTOCOL USING PURPLE EASYSEP® MAGNET (CATALOG #18000)This procedure is used for processing **500 µL - 2 mL** of sample (up to 1×10^8 cells).1. Prepare nucleated cell suspension at a concentration of 5×10^7 cells/mL in recommended medium (see Notes and Tips, reverse side). Cells must be placed in a 5 mL (12 x 75 mm) polystyrene tube to properly fit into the Purple EasySep® Magnet.*Falcon™ 5 mL Polystyrene Round-Bottom Tubes (BD Biosciences, Catalog #352058) are recommended.*2. Add EasySep® Human B Cell Enrichment Cocktail at **50 µL/mL cells** (e.g. for 2 mL of cells, add 100 µL of cocktail). Mix well and incubate at room temperature (15 - 25°C) for **10 minutes**.

3. Vortex EasySep® D Magnetic Particles for 30 seconds. Ensure that the particles are in a uniform suspension with no visible aggregates.

4. Add the D Particles at **75 µL/mL cells** (e.g. for 2 mL of cells, add 150 µL of particles). Mix well and incubate at room temperature (15 - 25°C) for **5 minutes**.5. Bring the cell suspension up to a **total volume** of 2.5 mL by adding recommended medium. Mix the cells in the tube by gently pipetting up and down 2 - 3 times. Place the tube (without cap) into the magnet. Set aside for **5 minutes**.6. Pick up the EasySep® Magnet, and in one continuous motion invert the magnet and tube, pouring off the desired fraction into a new 5 mL polystyrene tube. The magnetically labeled unwanted cells will remain bound inside the original tube, held by the magnetic field of the EasySep® magnet. Leave the magnet and the tube in inverted position for 2 - 3 seconds, then return to upright position. *Do not shake or blot off any drops that may remain hanging from the mouth of the tube.* The negatively selected, enriched cells in the new tube are now ready for use.**C) MANUAL EASYSEP® PROTOCOL USING "THE BIG EASY" SILVER EASYSEP® MAGNET (CATALOG #18001)**This procedure is used for processing **500 µL - 8.5 mL** of sample (up to 4.25×10^8 cells).1. Prepare nucleated cell suspension at a concentration of 5×10^7 cells/mL in recommended medium (see Notes and Tips, reverse side). Cells must be placed in a 14 mL (17 x 100 mm) polystyrene tube to properly fit into the Silver EasySep® Magnet.*Falcon™ 14 mL Polystyrene Round-Bottom Tubes (BD Biosciences, Catalog #352057) are recommended.*2. Add EasySep® Human B Cell Enrichment Cocktail at **50 µL/mL cells** (e.g. for 2 mL of cells, add 100 µL of cocktail). Mix well and incubate at room temperature (15 - 25°C) for **10 minutes**.

3. Vortex EasySep® D Magnetic Particles for 30 seconds. Ensure that the particles are in a uniform suspension with no visible aggregates.

4. Add the D Particles at **75 µL/mL cells** (e.g. for 2 mL of cells, add 150 µL of particles). Mix well and incubate at room temperature (15 - 25°C) for **5 minutes**.5. Bring the cell suspension to a **total volume** of 5.0 mL (for $< 2 \times 10^8$ cells) or 10 mL (for $2 - 4.25 \times 10^8$ cells) by adding recommended medium. Mix the cells in the tube by gently pipetting up and down 2 - 3 times. Place the tube (without cap) into the magnet. Set aside for **5 minutes**.6. Pick up the EasySep® Magnet, and in one continuous motion invert the magnet and tube, pouring off the desired fraction into a new 14 mL polystyrene tube. The magnetically labeled unwanted cells will remain bound inside the original tube, held by the magnetic field of the EasySep® magnet. Leave the magnet and the tube inverted for 2 - 3 seconds, then return to upright position. *Do not shake or blot off any drops that may remain hanging from the mouth of the tube.* The negatively selected, enriched cells in the new tube are now ready for use.

Components:

• EasySep® Human B Cell Enrichment Cocktail	5.0 mL
• EasySep® D Magnetic Particles	2 x 5.0 mL

REQUIRED EQUIPMENT:

EasySep® Magnet (Catalog #18000), or the "Big Easy" EasySep® Magnet (Catalog #18001), or "Easy 50" EasySep® magnet (Catalog #18002).

PRODUCT DESCRIPTION AND APPLICATIONS:

EasySep® Human B Cell Enrichment Cocktail and EasySep® D Magnetic Particles label non-B cells for magnetic separation. These reagents are designed to enrich B cells from normal fresh or previously frozen peripheral blood mononuclear or nucleated cells by depletion of non-B cells.

EASYSEP® LABELING OF HUMAN CELLS:

Unwanted cells are specifically labeled with dextran-coated magnetic particles using bispecific Tetrameric Antibody Complexes (TAC). These complexes recognize both dextran and the cell surface antigen expressed on the unwanted cells (Figure 1). The small size of the magnetic dextran iron particles allows for efficient binding to the TAC-labeled cells. Magnetically labeled cells are then separated from unlabeled target cells using the EasySep® procedure (reverse side).




Figure 1.
Schematic Drawing of EasySep® TAC Magnetic Labeling of Human Cells.

NOTES AND TIPS:

PREPARING A NUCLEATED CELL SUSPENSION

FROM WHOLE PERIPHERAL BLOOD

Prepare a mononuclear cell suspension from whole peripheral blood by Ficoll-Paque™ PLUS density separation (Catalog #07957). For previously frozen mononuclear cells, we recommend incubating the cells with DNase I (Catalog #07900) at a concentration of 100 µg/mL for at least 15 minutes at room temperature (15 - 25°C) prior to labeling and separation. Filter clumpy suspensions through a 30 µm mesh nylon strainer for optimal results. Resuspend cells at 5×10^7 cells/mL in recommended medium.

FROM PERIPHERAL BLOOD APHERESIS (LEUKOPAK)

If working with large volumes (>150 mL), concentrate Leukopak cells first by centrifuging at 500 x g for 10 minutes, removing the supernatant, and resuspending the cells in 1/10th of the original Leukopak volume in recommended medium (e.g. for 300 mL of cells, resuspend in 30 mL of recommended medium). For small volumes (150 mL or less), add the ammonium chloride solution directly to the cell suspension.

1. Add an equal volume of Ammonium Chloride solution to the Leukopak suspension (e.g. for 5 mL of Leukopak suspension, add 5 mL Ammonium Chloride solution).
2. Incubate 15 minutes on ice.
3. Centrifuge at 500 x g for 10 minutes at room temperature. Remove the supernatant.
4. Wash the cells by topping up the tube with the recommended medium. Centrifuge the cells at 150 x g for 10 minutes at room temperature with the brake off. Carefully remove the supernatant.
5. Repeat the wash step one or more times until most of the platelets have been removed (indicated by a clear supernatant).
6. Resuspend cells at 5×10^7 cells/mL in the recommended medium.

OPTIMAL CELL NUMBER. The use of fewer than 5×10^7 cells per separation may result in sub-optimal performance.



NEGATIVE SELECTION


NOTES AND TIPS CONTINUED:

RECOMMENDED MEDIUM. The recommended medium is RoboSep® buffer (Catalog #20104) or Phosphate Buffered Saline (PBS) + 2% FBS (Catalog #07905) with 1 mM EDTA added. Medium should be Ca⁺⁺ and Mg⁺⁺ free.

ASSESSING PURITY. Purity of B cells can be measured by flow cytometry after staining with a fluorochrome-conjugated antibody against CD19 (e.g. FITC anti-CD19, Catalog #10409), CD20 or other B cell marker.

TYPICAL EASYSEP® B CELL ENRICHMENT PROFILE:

Start: 8% CD19⁺ Cells Enriched: 99% CD19⁺ Cells



Starting with mononuclear cells, the CD19⁺ cell content of the enriched fraction typically ranges from 95 - 99%.

COMPONENT DESCRIPTIONS:

EASYSEP® HUMAN B CELL ENRICHMENT COCKTAIL

CODE #19054CVX

This cocktail contains a combination of monoclonal antibodies purified from hybridoma culture supernatant by affinity chromatography using Protein A or Protein G Sepharose. These antibodies are bound in bispecific Tetrameric Antibody Complexes (TAC) which are directed against cell surface antigens on human blood cells (CD2, CD3, CD14, CD16, CD36, CD43, CD56, CD66b, glycoporphin A) and dextran. The mouse monoclonal antibody subclass is IgG1. It should be noted that this product is a biological reagent, and as such cannot be completely characterized or quantified. Some variability is unavoidable.

EASYSEP® D MAGNETIC PARTICLES

CODE #19250VX

A suspension of magnetic dextran iron particles in TRIS buffer.

STABILITY AND STORAGE:

EASYSEP® HUMAN B CELL ENRICHMENT COCKTAIL

Product stable at 2 - 8°C until expiry date as indicated on label. Contents have been sterility tested. Do not freeze this product. This product may be shipped at room temperature (15 - 25°C), and should be refrigerated upon receipt.

EASYSEP® D MAGNETIC PARTICLES

Product stable at 2 - 8°C until expiry date as indicated on label. Contents have been sterility tested. Do not freeze this product. This product may be shipped at room temperature (15 - 25°C), and should be refrigerated upon receipt.

Ficoll™ and Ficoll-Paque™ PLUS are trademarks of GE Healthcare Ltd.