

Meniscus-free cultureware for more accurate counting of hematopoietic colonies in CFU assays

Catalog # 27370	5/Pack
27371	50/Pack

Product Description

SmartDish™ 6-well culture plates enable accurate and reproducible hematopoietic colony counting by preventing the formation of a meniscus. This allows for an even distribution of culture medium, resulting in a more uniform distribution of colonies throughout the entire 35 mm well. The absence of the meniscus also reduces optical distortion so that colonies located at the edge can be more easily counted.

SmartDish™ (Catalog #27370) is the second generation of the original SmartDish™ plate (Catalog #27301). These new plates use an improved coating, which streamlines the manufacturing process while consistently improving the visibility of hematopoietic colonies in culture through elimination of the meniscus.

SmartDish™ cultureware has been designed to work with standard inverted microscopes when performing manual colony counts, as well as with the STEMvision™ instrument (Catalog #22000) for automated counting of hematopoietic colony-forming unit (CFU) assays. For manual counting of hematopoietic CFU assays, SmartDish™ is used with STEMgrid™-6 (Catalog #27000), a detachable grid that facilitates navigation throughout the culture, dividing it into four quadrants for partial counting if desired.

Properties

Storage: Store at 15 - 25°C.

Shelf Life: Not applicable

Handling / Directions For Use

SmartDish™ meniscus-free cultureware is intended for use in the same manner as conventional cultureware. In order to ensure that cultures in SmartDish™ cultureware are in a well-hydrated environment, add 4 mL of sterile water to the empty spaces between the SmartDish™ wells. It is also recommended that at least four additional 35 mm culture dishes (e.g. Catalog #27100), each containing 3 - 4 mL of sterile water, are placed around the SmartDish™ cultureware and that all are contained within a larger tray, e.g. a 245 mm x 245 mm Square Treated Tissue Culture Dish (Catalog #27140).

NOTE: Up to three SmartDish™ 6-well plates can fit in a 245 mm x 245 mm Square Treated Tissue Culture Dish.

For complete instructions on how to use SmartDish™ 6-well plates with MethoCult™ media, refer to the Technical Manual: Human Colony-Forming Unit (CFU) Assays Using MethoCult™ (Document #28404), available at www.stemcell.com or contact us to request a copy.

For complete instructions on how to use SmartDish™ 6-well plates with STEMvision™, refer to the Technical Manual: STEMvision™ Automated Colony-Forming Unit (CFU) Assay Reader (Document #29973), available at www.stemcell.com or contact us to request a copy.

STEMCELL TECHNOLOGIES INC.'S QUALITY MANAGEMENT SYSTEM IS CERTIFIED TO ISO 13485. PRODUCTS ARE FOR RESEARCH USE ONLY AND NOT INTENDED FOR HUMAN OR ANIMAL DIAGNOSTIC OR THERAPEUTIC USES UNLESS OTHERWISE STATED.

Copyright © 2017 by STEMCELL Technologies Inc. All rights reserved including graphics and images. STEMCELL Technologies & Design, STEMCELL Shield Design, Scientists Helping Scientists, MethoCult, SmartDish, STEMgrid, and STEMvision are trademarks of STEMCELL Technologies Canada Inc. All other trademarks are the property of their respective holders. While STEMCELL has made all reasonable efforts to ensure that the information provided by STEMCELL and its suppliers is correct, it makes no warranties or representations as to the accuracy or completeness of such information.